Specifications

Model name	Handy Hardness Tester SONOHARD	Display	a. Measured value: 3 digits		
	(Motorized/manual switchover type p	orobe)	make-up	b. TIMES: 2 digits	
Measuring indenter	Diamond indenter for Micro-Vickers			(measuring frequency)	
	(facing to surface angle of 136°)		c. MAX value: 3 digits		
Test load and control no.	1. Approx 2kgf (roughly 20N)		d. MIN value: 3 digits		
CE making complied models are	Z. Approx Tkgt (roughly TUIN)	SH-21-E1		e. σ: 4 digits	
required to add (CE) with control no.	3. Approx 4kgf (roughly 40N)	SH-21-E4		(standard deviation)	
	(special type manual probe only)			f. \overline{x} : 4 digits (average value)	
Measuring range	1. Rockwell hardness HRC 10.0~7				
	2. Vickers hardness HV 100~999)		,	
	3. Shore hardness HS 20.0∼99.	.9			
	4. Brinell hardness HB 85~550				
Reproducibility	HRC: \pm 1.0HRC, HV: \pm (3%rdg)H	V,	Set-up	a. UPPER	
	HS: \pm 1.0HS, HB: \pm (3%rdg)H	lB		(upper limit) b. LOWER (lower limit)	
Applicable test materials	With steel as the principle material, of	other metallic			
	materials may also be measured by	calibrating			
	against a standard hardness test blo		c. TIMES		
Display of measured values	Digital display (LCD, 4 digits) with E		(measuring frequency)		
Data memory	2,000 pieces		d. CANCEL		
Digital display units	1HV, 0.1HRC, 0.1HS, 1HB, 1N/mm	Alarm	a. Alarm (buzzer sound)		
Allowable operating temperature	0~50°C	Output	a. RS-232C output		
Power source	AC adapter(100~240V) or lithium id	Ουιραί	used for data		
Continious operating hours	5 hours when featuring a rechargeable		transmission or printing		
	*BL= backlight			transmission or printing	
Dimensions	Display unit: 97(W) x 50(D) x 170(H	I) mm,	Frequency	1. Motorized / manual	
	Probe diameter: 40mm, length: 164mm		rioquorioy	switchover type approx 69~71kHz 2. Manual type (special type)	
Weight					
Carrying case dimensions	400(W) x 220(D) x 140(H) mm			approx 38~40kHz	
Standard components	1 display unit, 1 probe (including standard attachment/UA5410)			HV、HRC、HS、HB、N/mm² according to JIS B 7731、 ASTM E 140、DIN 50150	
	1 hardness standard test block (arround HRC55),				
	1 probe cable 1.5m, 1 AC adapter 100~240V (A05C1-09MI),				
	1 recharger (MK-8220), 1 lithium ior				
	1 carrying case (MK-9701)				
Options	Standard hardness test blocks HV60	00, HS50, HB300 used f	or scale calil	oration,	
	Measuring stand for small objects (SH-S02), Probe attachment for pipes materials (UA4-4005), Probe attachment for inner races (UA4-4006), Printer model DPUH245AS with cable,				
	Printer paper (TP-H241L), Stand for the main unit (SH-P03				

- When using the tester installed in automated machinery, please contact our hardness tester sales department for specifications concerning the testers used for automatic machines.
- In order to make it possible to propose ideas regarding the unique loads, jigs and measuring methods to our customers. we humbly request that you fill out a specification confirmation form when making inquiries concerning usage
- The SONOHARD SH-21(E) is calibrated using the standard hardness test block produced in compliance with JIS B7730/ ISO 6508-3 and JIS B7735/ ISO 6507-3 by Yamamoto Scientific Tool Laboratory Co., Ltd., Japan, who has the quality management system approved under ISO 9001. The values measured by SH-21(E) are therefore guaranteed by us. (Measured values under the calibration using other makers' test blocks are out of our guarantee.)
- The model name on the catalog is SH-21(E), while it is referred to as SH-21 only in the relevant operation manual,
- A standard export model of SH-21(E) is not CE-Marking complied, but a CE-Marking complied model is also available by factory modification on
 - * Please read the user's manual before undertaking operations

12.08.data. 🔘



Tokyo Head Office (Overseas Sales)


FE Kuramae Bldg., 2-17-4 Kuramae, Taito-ku, Tokyo

e-mail: tokyo@jfe-advantech.co.jp Tel. +81-3-5825-5577 Fax. +81-3-5825-5591


e-mail: honsha@ife-advantech.co.ip Tel. +81-798-66-1508 Fax. +81-798-65-7025

3-48. Takahata-cho, Nishinomiya-shi, Hyogo Pref.

Handy Hardness Tester




Perfect for use in making on-site measurements $oldsymbol{\xi}$






The Handy Hardness Tester (SONOHARD®) model SH-21(E) differs completely from traditional hardness testers from a viewpoint that instead of measuring the size of the indentation of the test sample using a microscope, it employs a diamond indenter equipped with a vibrating rod that presses on the test surface at a fixed force and then measures its hardness by applying ultrasonic vibrations.

When the vibrating rod is applied to a soft-surfaced test sample with identical qualities and at a fixed force, it makes a deep indentation and gets locked into the groove. Due to this, the resonance frequency increases. Conversely, it does not get locked in when used on hard test samples and the resonance frequency drops. The test sample's hardness can be calculated using the correlation between this deviation and the tested hardness.





Calculation values of SH indentation

Load P= approx. 2kgf (approx. 20N)

Hardness	Calculation value,	Calculation value,	Conversion value,
HV	Size of indentation (mm)	Depth of indentation (mm)	HRC
100	0.193	0.028	_
200	0.136	0.019	15
300	0.111	0.016	31
400	0.096	0.014	41
500	0.086	0.012	49
600	0.079	0.011	55.5
700	0.073	0.010	60.5
800	0.068	0.010	64.5
900	0.064	0.009	67.5

Load P= approx. 1kgf (approx. 10N)

Hardness HV	Calculation value, Size of indentation (mm)	Calculation value, Depth of indentation (mm)	Conversion value, HRC
100	0.136	0.019	_
200	0.096	0.014	15
300	0.079	0.011	31
400	0.068	0.010	41
500	0.061	0.009	49
600	0.056	0.008	55.5
700	0.051	0.007	60.5
800	0.048	0.007	64.5
900	0.045	0.006	67.5

Probe dimensions unit / mm 100.0 (3) Vibrating rod ø40 Cover ø45 ø50

Precautions on measurements

1. The affect of surface roughness

	x:Average value		e σ:Standard dev	iation Measuremen	Measurement frequency per 100	
	Hardness	Surface roughness	0.8a	1.6a	3.2a	
HRC31	LIDO04 E	x	31.5	31.7	30.9	
	HRC31.5	σ	0.4	0.5	0.8	
HRC	110050.0	X	50.5	50.5	50.3	
	HRC50.8	σ	0.3	0.3	0.6	
HRC65.5		x	65.4	65.3	65.1	
	HRC65.5	σ	0.2	0.2	0.4	

• For items with a roughness of 3.2a or greater, you will need to polish the surface before

2. Measurable dimensions (For loading of 2kgf/approx. 20N)

1 Size : 15mm wide x 15mm long or greater 2 Thickness t= 7mm or greater • For items smaller in dimension than those listed above, you will need to use a stand with dimensions of \$60 x 15mm or greater, and make measurements after securing the item to the stand by applying oil and after increasing the virtual mass.

● For loading of 1kgf/approx.10N, it is possible to make measurements even at t=7mm or less.

Make measurements within ±3.

0° 1.9° 3.8° 5.7°

4. Reproducibility When making measurements on a standard test block using a measuring stand

5.7° 3.8° 1.9°


3. Angles and deviations

		Standard block	Stand measurement	Hand-held measurement
	HV Average hardness value	HV200	199	201
		500	504	508
		800	806	790
				¾ 30 measurements

Utilization of the Handy Hardness tester SH-21

Examples of quality control and maintenance usage by measuring handness







Tapered parts



Car wheels



Diagnosing wear and tear using



Drill blade



Measuring the strength of welding sections (Checking tension strength)

Measuring metal fatigue in steel towers, bridges and reinforcing bars







Measurements with Measuring stand for small objects

