



VERTICAL AUTOCLAVE

**Ideal for use in Medical clinics, Health centers,
Hospital departments, Dental clinics, Acupuncture
clinics, Laboratories...**

With Drying Cycle



CE 0434
ISO 9001
ISO 13485

RAU-530D **Designed For Better Life**



REXMED Industries Co., Ltd.

P.O.BOX 487, Kaohsiung, Taiwan

Tel: +886-7-3814010 Fax: +886-7-3814013

E-mail: info@rexmed.com, sales@rexmed.com

http://www.rexmed.com



RAU-530D




 Cosmos Supply Co., Ltd.
 cosmos_supply@hotmail.com
 cosmos_supply@yahoo.co.th

 QTS Agro Co., Ltd.
 qts_agro@hotmail.com
 qts_agro@yahoo.co.th
www.cosmos-supply.com

STERILIZATION TIME AND TEMPERATURE						
Water Supply	Air Evacuation	Pressure Build Up Time	Sterilization Time	Hot Water & Steam Exhaust	Drying Time	Necessary Time
1.5 liters	19 min	8 min	20 min	1 min	60 min	Total 78 min

RECOMMENDED PERIODS OF STERILIZE EXPOSURE 121°C	
Material to be sterilized	Time in minutes
Instruments in tray	20 min
Syringes & needles	20 min
Rubber gloves	20 min
Fabrics woven	30 min
Solutions	30 min

STANDARD ACCESSORIES	
Wire basket	2pcs
Heating element	1pc
Silicone door seal	1pc

SPECIFICATIONS			
Voltage	AC 110/230V	Auto add water	Manual
Power supply	2 KW	Chamber size	30 x 71 cm SS#304
Sterilizing pressure	1.2~2.1 kg/cm ²	Capacity	50 Liters
Sterilizing	121~134°C	N.W.	70 KGS
Auto dry	Yes	Packing size	L60 X W47 X H112 cm

REXMED Designed For Better Life